



Mr. Dindo Carillo  
Official FDA Correspondent  
Teco Diagnostics  
4925 E. Hunter Avenue  
Anaheim, California 92807

**MAR 10 2000**

Re: X000001  
Regulation Number: 862.2900  
Product Code: KQO

### **Categorization Notification**

Regulations codified at 42 CFR 493.15 et seq., implementing the Clinical Laboratory Improvement Amendments of 1988, require the Secretary to provide for the categorization of specific clinical laboratory test systems by the level of complexity. Based upon these regulations, the following commercially marketed test system or assay for the analyte is categorized below:

Test System: Teco Diagnostics URITEK TC-101 Urine Strip Reader  
Analyte: Urine Qualitative Dipstick Chemistries                      Complexity: Waived

This complexity categorization is effective as of the date of this notification. This categorization will be reported on FDA's home page <http://www.fda.gov/cdrh/clia>. This categorization information may be provided to the user of the commercially marketed test system or assay as specified for the analyte(s) indicated. This categorization will also be announced in a Federal Register Notice, which will provide opportunity for comment on the decision. FDA reserves the right to reevaluate and recategorize this test based upon the comments received in response to the Federal Register Notice. For questions regarding FDA's categorization procedure, contact Clara A. Sliva, Acting CLIA Coordinator, at ((301) 827-0496 or email <http://www.CLIA@CDRH.FDA.GOV>.

Sincerely yours,

Steven I. Gutman, MD  
Director  
Division of Clinical Laboratory  
Devices  
Office of Device Evaluation  
Center for Devices and Radiological  
Health