

Part Number Configurations

Connex® Vital Signs Monitor					
6	X	X	X	X	X
Model	Radio	Pulse Oximetry	Respiration	Thermometry	Printer
6	7 = Not Included	N = Nellcor® SpO ₂	C = CO ₂	T = SureTemp® Plus	P = Internal
	8 = Included	M = Masimo® SpO ₂	R = RRa	E = Braun 6000	X = Not Included
		H = Masimo SpO ₂ and Total Hemoglobin (SpHb)	S = EarlySense® F = ES Safety	D = Both	
			X = Not Included	X = None	

Note: SpHb and RRa cannot be configured together.

Connex® Integrated Wall System						
8	X	X	X	X	X	X
Model	Radio	Pulse Oximetry	Oral Temp	Storage	Ear Temp	Heads
8	4 = Not Included	N = Nellcor SpO ₂	T = SureTemp Plus	V = SCS (Standard Cuff Storage)	E = Braun 6000	C = MacroView™ 23820 + Coaxial Ophthalmoscope 11720
	5 = Included	M = Masimo SpO ₂	X = No Oral Temp		X = No Ear Temp	P = MacroView 23820 + PanOptic™ Ophthalmoscope 11810
		X = No SpO ₂				X = No Heads

For Connex devices, power cords are indicated as “-X” at the end of the part number as a 1-digit alphanumeric code (e.g., 84MTVEC-B or 67MXTP-B).

Power Cord Key	
B	North America
2	Europe
4	United Kingdom
6	Australia/New Zealand
7	South Africa

Model	67	68	84	85
Wireless Connectivity				
Upgradeable	✓		✓	
Integrated Radio		✓		✓
Wired Connectivity				
USB	✓	✓	✓	✓
Ethernet	✓	✓	✓	✓
Nurse Call	✓	✓	✓	✓
Number of Accessories Capable of Running	4	4	4	4